

Certificate of Compliance

Catalog Numbers:

Non-sterile Products: _____

Sterile Products: _____ AR-1250-SLF _____

Product Description: 1000µl Filter universal pipette tips-scale ring marks , Clear Racked, Pre-sterilizeLow Retention

Mfg. Lot Number: 201120-1138

Material: Virgin Polypropylene

Sterilization:

Sterile catalog numbers from this lot have been pre-sterilized and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*. Sterility Assurance Level: SAL 10⁻⁶

This does not apply to non-sterile product catalog numbers.

Endotoxin (pyrogen) Testing:

Products are directly tested for endotoxins by the Limulus Ameobacyte Lysate (LAL) gel clot assay according to USP/FDA guidelines for medical devices. No pyrogens were detected.

Results: Test sample extraction had an undetectable level of endotoxin at **0.05 EU/mL**

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Lot Number Designation:

The product lot number shall be 6 digits followed by a "-" and 4 digits. The first two digits of the lot number shall be the last two digits of the current year. The next two digits shall represent the current month. The last two of the six digits shall represent the current date of the month. The last three numbers are preceded by a "-" and represents the actual mold number.

Oxford Lab Products's practice is to recommend product use within three years from the date of manufacture. This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released for the following characteristics:

Test/Procedure	Result
Visual attributes	Pass
Dimensional	Pass
Accuracy Reading	Pass
Tip warp	Pass

Quality Manager: Weiming Wan

Date: November 26, 2020