

Quality Certificate



Product Name: 100µL Pipette Micro Tips

Catalog Number: OAR-100-SLF

Batch Number: 210103-305

Expiry Date: 20240103

Customer Focus:

At Oxford Lab Products, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

Satisfaction Guarantee:

Oxford Lab Products offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Oxford, we will either replace or issue a refund for the purchase price of your product.

Oxford Lab Products are manufactured in accordance with ISO 13485: 2016 & ISO 9001: 2015 quality management systems in a class 100,000 clean room.

Sterilization - Product labeled as sterile is irradiated and dosimetrically released upon ISO 11137 recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. Oxford Lab Products labeled sterile meet a minimum requirement of 10^{-6} SAL (Sterility Assurance Level).

Low Retention Tip - Product is treated by special surface treatment process that makes the tip has super-hydrophobic characteristic and leaves barely residues in the body of our tips. This will not waste precious samples, and ensure accuracy of the experimental data.

Non-Pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate testing for medical devices and company guidelines. The acceptance level for product is less than 0.5 EU/mL.

DNase & RNase Free - This product has been tested and is free of any detectable DNase / RNase contamination.

Nucleic Acid - This product has been tested and is free of any detectable Nucleic Acid.

Trace Metal - All material are certified metal-free by material manufacturers before they are turned into production which compliance with US Pharmacopoeia guidelines.

ATP free - This product has been tested and is free of ATP.

PCR inhibitor - Products are tested via PCR amplification and gel electrophoresis analysis. Samples must show normal amplification to be considered free of PCR inhibitors.

Quality Control Testing - This product has been inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

A handwritten signature in black ink, appearing to read "York Zhang", is positioned above a horizontal line.

Quality Manager
York Zhang